



Product Code: LVA#911NW-S

Description: Sterile 70% USP IPA 30% WFI Non-woven Polyester/Cellulose Wipes

Quantity in 1 case: Flat pack 20 wipes/pouch; 24 pouches/case

Lot#: 03000221

Expiration Date: Apr 2023

## CERTIFICATE OF ANALYSIS

<u>Parameters</u>	<u>Specification</u>	<u>Result</u>
Basis Weight	68 ± 4 g/m <sup>2</sup>	64 g/m <sup>2</sup>
Thickness	0.27 – 0.34 mm	0.30 mm
Absorbency Capacity ( <i>TEST-RP-CC 004.3</i> )	≥ 280mL/m <sup>2</sup>	289mL/m <sup>2</sup>
Absorbency Rate ( <i>TEST-RP-CC 004.3</i> )	≤ 0.5 sec	0.5 sec
Colour	Colourless	Colourless
Odour	Alcohol – like	Alcohol – like
Solution Clarity	Clear	Clear
IPA Content	68.0 to 72.0 %	69.06%
Specific Gravity at 20°C	0.872 to 0.883	0.8731
Refractometer	24 – 26	25.5
Sterility Test ( <i>ISO 11137 method</i> )	No growth observed	No growth observed

### Irradiation Details

**Date: 23 Apr 2021**

Specified Dose (kGy)

Delivered Dose (kGy)

Min: 25      Max: 45

Min: 26.2      Max: 29.1

*This product / product category has been validated for Sterility Assurance Level of SAL 10<sup>-6</sup> with VD Max<sup>25</sup> Method as per ISO EN 11137-2:2015: Sterilization of healthcare products – Radiation Part 2 – Establishing the Sterilization Dose; with a conclusion of a dose of 25kGy will provide Sterility Assurance Level of SAL 10<sup>-6</sup>.*

Reviewed and Approved by:

Date:

28 Apr 2021

For more information, contact Fisher Scientific

**Singapore:**

For customer service, call +65 68734006  
Email: enquiry.sg@thermofisher.com  
fisher.sg | myfisherstore.com/singapore

**Malaysia:**

For customer service, call +603 5122 8888  
Email: enquiry.my@thermofisher.com  
fisher.my | myfisherstore.com/malaysia

For Research Use Only. Not for use in diagnostic procedures. © 2017 Thermo Fisher Scientific Inc. All rights reserved.  
All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.

