

Product Code: LVA# 7031 Description: Sterile 70% IPA, 30% WFI 946ml Bag-in-Bottle Lot#: 30000223 Expiration Date: 28 Feb 2026

CERTIFICATE OF ANALYSIS

<u>Parameters</u> Colour	Specification Colourless	<u>Result</u> Colourless
Odour	Alcohol – like	Alcohol – like
Clarity	Clear	Clear
Filtered	To 0.2 microns	Using 0.2 micron
IPA Content	68.0 to 72.0 %	68.87%
Specific Gravity at 20°C (kg/m ³)	0.872 to 0.883	0.8730
Refractometer	24 – 26	25.5
Bacteria Endotoxin Test (LAL kit by Cape Cod)	< 0.25 EU/ml	< 0.25 EU/ml
Sterility Test (ISO 11137 method)	No growth observed	No growth observed

Irradiation Details Specified Dose (kGy)

Min: 25 Max: 45

'This product / product category has been validated for Sterility Assurance Level of SAL 10⁻⁶ with VD Max²⁵ Method as per ISO EN 11137-2:2015: Sterilization of healthcare products – Radiation Part 2 – Establishing the Sterilization Dose; with a conclusion of a dose of 25kGy will provide Sterility Assurance Level of SAL 10^{-6'}.

Reviewed and Approved by:

For more information, contact Fisher Scientific

Singapore: For customer service, call +65 6873-6006

Malaysia: For customer service, call +603 5122 8888 Tanil and universities and the communication of the second second

For Research lies Only. Not for use in diagnostic procedures. © 2017 Tharmo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.

Date:

Min: 27.8

Delivered Dose (kGy)

Max: 31.6

07 Mar 2023

