

Product Code: LVA#7032 Description: Sterile 70% IPA, 30% WFI (Screw Cap Bottles) Quantity in 1 case: 4 x 1 gallon/case Lot#: 30000622 Expiration Date: 30 Sep 2025

## **CERTIFICATE OF ANALYSIS**

<u>Parameters</u> Colour	Specification Colourless	<u>Result</u> Colourless
Odour	Alcohol – like	Alcohol – like
Clarity	Clear	Clear
Filtered	To 0.2 microns	Using 0.2 microns
IPA Content	68.0 to 72.0 %	69.94%
Specific Gravity at 20°C (kg/m <sup>3</sup> )	0.872 to 0.883	0.8724
Refractometer	24 – 26	25.3
Bacteria Endotoxin Test (LAL kit by Cape Cod)	< 0.25 EU/ml	< 0.25 EU/ml
Sterility Test (ISO 11137 method)	No growth observed	No growth observed

## **Irradiation Details** Specified Dose (kGy)

Min: 25 Max: 45

**Delivered Dose (kGy)** Min: 28.8 Max: 31.1

'This product / product category has been validated for Sterility Assurance Level of SAL 10<sup>-6</sup> with VD Max<sup>25</sup> Method as per ISO EN 11137-2:2015: Sterilization of healthcare products – Radiation Part 2 – Establishing the Sterilization Dose; with a conclusion of a dose of 25kGy will provide Sterility Assurance Level of SAL 10<sup>-6</sup>'.

Reviewed and Approved by:

For more information, contact Fisher Scientific

Singapore: For customer service, call +65 6873-6006 Email, enquiry.sg@thermofisher.com

Malaysia: For customer service, call +603 5122 8888 Email, enquiry.my@thermofisher.com fishersci.com.sg | myfisherstore.com/singapore fishersci.com.my | myfisherstore.com/malaysia

For Research Use Only. Not for use in disgneetie procedures. © 2017 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.

Date:

03 Feb 2023

